

Comparative evaluation of ion release from orthodontic mini-implants in 2 mouthwashes: An in-vitro study

Mohammed Mandsaurwala^{1,*}, Ajit Kalia², Gaurav Gupta³, Ashwith Hegde⁴, Nasim Mirdehghan⁵

¹Postgraduate Student, ²Head of Department, ^{3,4}Reader, ⁵Senior Lecturer, Department of Orthodontics, M. A. Rangoonwala Dental College Pune.

***Corresponding Author:**

E-mail: mohammed_kutub@rediffmail.com / mohammedmandsaurwala@gmail.com

ABSTRACT:

Introduction: Orthodontic mini-implants can release metal ions into the saliva. NiTi orthodontic mini-implants from different manufacturers would have different corrosion resistance. Fluoridated mouthwashes are often recommended to orthodontic patients to reduce the risk of white-spot lesions around their mini-implants. We assayed the corrosion resistance, in terms of ion release, of different NiTi orthodontic mini-implants in different mouthwashes. The purpose of this study is to measure the levels of metal ions released from orthodontic mini-implants after immersion in different mouthwashes.

Methods: 30 orthodontic mini-implants from two different companies each, divided randomly into 3 equal groups were immersed in HEXIDINE (ICPA health products, India), LISTERINE mouthwash (Johnson & Johnson Ltd. Bangalore, India), and distilled deionized water incubated at 37°C for 30 days. Nickel, chromium, iron, copper, and manganese released from the orthodontic mini-implants were measured with an inductively coupled plasma spectrometer.

Results: The results showed that ion release in Listerine mouthwash was significantly ($P > 0.05$) higher than in the hexidine mouthwash and distilled water group.

Conclusion: CHX-containing (HEXIDINE) mouthwash can be offered to patients who have orthodontic appliances rather than Na + alcohol-containing mouthwash (LISTERINE).

Keywords: Ion Release, Mini-Implants, Mouthwash.

Access this article online	
Quick Response Code:	Website: www.innovativepublication.com
	DOI: 10.5958/2395-499X.2015.00007.6

INTRODUCTION

Anchorage control is an important factor in the success of orthodontic treatment. There have been many attempts to devise suitable anchorage methods, including intraoral and extraoral appliances. Mini-implants (MI) are anchorage devices that were introduced in orthodontics in the past two decades and whose use quickly spread in clinical practice.

The small size of mini-screw implants allows them to be placed into bone between the teeth, thus expanding their clinical applications⁽¹⁾. With more patients treated with screw implants as anchorage, their stability is gathering attention⁽²⁾. In recent years, it has been reported that mini-implants corrosion can occur in the oral environment. In an acidic environment and in the presence of fluoride ions, the corrosion resistance of certain materials, particularly titanium and titanium alloys, can deteriorate⁽³⁾.

The harmful effects of nickel, its carcinogenicity, allergenicity, and mutating substances have been systematically investigated at the cell, tissue, organ, and

organism levels. Approximately 10% of the general population has a hypersensitive reaction to nickel⁽⁴⁻⁶⁾.

During orthodontic treatment, practitioners recommend their patients to use mouthwashes, especially since most are adolescents who do not always follow a satisfactory oral-hygiene regimen and have a high risk of dental caries. Although fluoride ions in the prophylactic agents have been reported to cause corrosion and discoloration, little information is available regarding the effect of different mouthwashes in ion release of orthodontic mini-implants⁽⁷⁾.

The purpose of this study was to measure the levels of metal ions released from orthodontic mini-implants after immersion in two mouthwashes.

MATERIAL AND METHODS

Thirty orthodontic mini-implants from two different companies Dentos India Pvt. Ltd. and S.K Surgicals each were used for this study. All mini-implants were used in as-received condition. The mini-implants were divided randomly into 4 equal groups and immersed in HEXIDINE (ICPA health products, India), LISTERINE mouthwash (Johnson & Johnson Ltd. Bangalore, India), and distilled deionized water. These mouthwashes were chosen because of their commercial availability and identical methods of application. Each bracket was incubated in an oven set at a constant temperature of 37°C in individual 20-mL plastic-capped vials containing 15 mL of 1 mouthwash solution or distilled deionized water for 45 days.

At the end of incubation for 45 days, the immersion solution was tested with an inductively coupled plasma (ICP) spectrometer. ICP was used in this study as it has the advantage of extracting each ion simultaneously and detecting the metals without the interference of other ions. Standard stock solutions (100 mg mL⁻¹) of chromium, copper, iron, manganese, and nickel were prepared by dissolving their appropriate salts in distilled deionized water. Each solution was analyzed for chromium, copper, iron, manganese, and nickel ions.

STATISTICAL ANALYSIS

One-way analysis of variance (ANOVA) was used to analyze the differences among mean ion concentrations in the 6 groups. The Post-Hoc Tukey’s test was used for multiple group comparisons. P-value<0.05 is considered to be statistically significant.


RESULTS

Intra group comparison of metals for Dentos India Pvt. Ltd. mini-implants showed that (Table1) the average level of Chromium did not differ significantly across three sub-groups (P-value>0.05 for all). The average level of Copper is significantly higher in Listerine compared to Hexidine and distilled water sub-groups (P-value <0.001 for both). The average level of Iron is significantly higher in Listerine compared to Hexidine and distilled water sub-groups (P-value<0.001 for both). The average level of Manganese is significantly higher in Listerine compared to Hexidine and distilled water sub-groups (P-value<0.001 for both). The average level of Nickel is significantly higher in Listerine and Hexidine compared to distilled water sub-groups (P-value<0.001 for both). The average level of Nickel is significantly higher in Hexidine compared to Listerine sub-group (P-value<0.001).

Table 1: Intra-group comparison of metals (Dentos Group, n = 30).

Metal (PPMx10 ⁻²)	Sub Group (Mouth wash)			P-values (Intra-Group Comparisons)		
	Listerine (n=10)	Hexidine (n=10)	Distilled water (n=10)	Listerine v Hexidine	Listerine v Distilled water	Hexidine v Distilled water
Chromium	0.0 ± 0.0	0.0 ± 0.0	0.0 ± 0.0	0.999 ^{NS}	0.999 ^{NS}	0.999 ^{NS}
Copper	7.95 ± 0.45	0.0 ± 0.0	0.0 ± 0.0	0.001 ^{***}	0.001 ^{***}	0.999 ^{NS}
Iron	9.96 ± 1.91	0.24 ± 0.12	0.0 ± 0.0	0.001 ^{***}	0.001 ^{***}	0.987 ^{NS}
Manganese	3.73 ± 0.29	0.0 ± 0.0	0.0 ± 0.0	0.001 ^{***}	0.001 ^{***}	0.999 ^{NS}
Nickel	9.77 ± 1.09	12.91 ± 0.26	0.32 ± 0.03	0.006 ^{**}	0.001 ^{***}	0.001 ^{***}

Values are Mean ± Standard error of mean (SEM). P-values by one-way analysis of variance (ANOVA) with Post-Hoc Tukey’s test for multiple group comparisons. P-value<0.05 is considered to be statistically significant. *P-value<0.05, **P-value<0.01, ***P-value<0.001.




Intra group comparison of metals for S.K.Surgical mini-implants showed that (Table2) The average level of Chromium did not differ significantly across three sub-groups (P-value>0.05 for all).The average level of Copper is significantly higher in Listerine compared to Hexidine and distilled water sub-groups (P-value<0.001 for both). The average level of Iron is significantly higher in Listerine compared to Hexidine and distilled water sub-groups (P-value<0.001 for both). The average level of Manganese is significantly higher in Listerine compared to Hexidine and distilled water sub-groups (P-value<0.05 for both). The average level of Nickel is significantly higher in Listerine and Hexidine compared to distilled water sub-groups (P-value<0.001 for both). The average level of Nickel is significantly higher in Hexidine compared to Listerine sub-group (P-value<0.001).

Table 2: Intra-group comparison of metals (Sk Group, n = 30).


Metal (PPMx10 ⁻²)	Sub Group (Mouth wash)			P-values (Intra-Group Comparisons)		
	Listerine (n=10)	Hexidine (n=10)	Distilled water (n=10)	Listerine v Hexidine	Listerine v Distilled water	Hexidine v Distilled water
Chromium	0.0 ± 0.0	0.08 ± 0.04	0.0 ± 0.0	0.999 ^{NS}	0.999 ^{NS}	0.999 ^{NS}
Copper	6.82 ± 1.64	0.0 ± 0.0	0.06 ± 0.02	0.001 ^{***}	0.001 ^{***}	0.999 ^{NS}
Iron	19.95 ± 9.79	0.03 ± 0.02	0.0 ± 0.0	0.049 [*]	0.048 [*]	0.999 ^{NS}
Manganese	2.12 ± 0.21	0.0 ± 0.0	0.0 ± 0.0	0.001 ^{***}	0.001 ^{***}	0.999 ^{NS}
Nickel	14.39 ± 1.03	12.16 ± 0.12	0.41 ± 0.03	0.113 ^{NS}	0.001 ^{***}	0.001 ^{***}

Values are Mean ± Standard error of mean (SEM). P-values by one-way analysis of variance (ANOVA) with Post-Hoc Tukey’s test for multiple group comparisons. P-value<0.05 is considered to be statistically significant. *P-value<0.05, **P-value<0.01, ***P-value<0.001.



Inter-group distribution and statistical comparison of metals showed that The average level of Chromium is significantly higher in Hexidine (Sk) group compared to Listerine (D), Hexidine (D) and Distilled water (D) sub-groups (P-value<0.05 for all).The average level of Copper is significantly higher in Listerine (D) group compared to Hexidine (S), Distilled water (S) sub groups (P-value<0.001 for all). The average level of Copper is significantly higher in Listerine (S) group compared to Hexidine (D), Distilled water (D) sub groups (P-value<0.001 for all). The average level of Iron is significantly higher in Listerine (S) compared to Hexidine (D) and distilled water (D) sub-groups (P-value<0.05 for both). The average level of Manganese is significantly higher in Listerine (S) compared to Listerine (D), Hexidine (D) and Distilled water (D) sub-groups (P-value<0.001 for all). The average level of Manganese is significantly higher in Listerine (D) compared to Hexidine (S) and Distilled water (S) sub-

groups (P-value<0.001 for all). The average level of Nickel is significantly higher in Listerine (S) compared to Listerine (D) and Distilled water (D) sub-group (P-value<0.001 for both). The average level of Nickel is significantly higher in Hexidine (S) compared to Distilled water (D) sub-group (P-value<0.001).



DISCUSSION

The incubation period of 45 days was decided assuming that each time the mouthwash was present for 6 hours in a patient's mouth (24 months, twice a week = about 69,000 minutes), so for this study the mini-implants were immersed in mouthwashes and incubated at 37°C for 45 days (45 days about 64,000 minutes)⁽⁸⁾. Also, several studies have demonstrated that the levels of metal release from fixed orthodontic appliances peak at day 7, and that all release is completed within 4 weeks. Although there are several studies that have investigated corrosion on orthodontic appliances caused by mouthwashes, there are not many studies that have investigated the effects of mouthwashes on different mini-implants⁽⁹⁻¹¹⁾.

Since for comparison of ion release in different mouthwashes, we needed a basic solution without influencing ions, we used deionized water to prevent the effects of saliva composition on the basic results. However, from our results, it can be concluded that the corrosiveness of the mouthwash, which in turn depends on its chemical structure, is the main factor responsible for the release of metal ions from dental mini-implants⁽⁹⁾.

Our finding is that the greatest amount of metal ion release occurs in Na + alcohol containing mouthwashes and may contribute to both the literature and to clinical

practice. Schiff et al. evaluated the corrosion effect of NaF containing mouthwashes with three different concentrations and concluded that 125 ppm NaF-containing mouthwashes should not be preferred for orthodontic appliances containing titanium, iron, chromium, and nickel⁽¹²⁾. The results of our study indicated that CHX-containing mouthwash caused lower amounts of metal release compared to other mouthwashes due to a greenish discoloration occurring on the CHX samples after 24 hours of immersion which constituted a type of passive surface. The pigments may be the reason for the lower metal ion release than expected⁽¹³⁾.

From a clinical viewpoint, the corrosion of mini-implants might affect their biocompatibility, and the final result of orthodontic treatment could be compromised⁽¹⁴⁾. Certain ions such as nickel and chromium can result in symptoms of toxicity and allergic reactions⁽¹⁵⁾. Although the reasons a miniscrew may induce inflammation are multifactorial and include patient hygiene level, type of surrounding tissue, and miniscrew head design⁽¹⁶⁾, it should be noted that corrosion has been implicated as one of several triggering factors associated with peri-implantitis of dental implants⁽¹⁷⁾. These symptoms can be short-lived and intense or longer lasting and moderate, and some might be resolved, whereas others can become a

chronic problem⁽¹⁸⁾. Also, severe gingivitis can be related not only to poor oral hygiene but also to a hypersensitivity reaction to nickel or chromium ions released from mini-implants^(19,20). We also need to determine whether these ion releases have clinical significance in sensitizing patients with a history of hypersensitivity.

Metal is released into the oral cavity with saliva as the medium, and this could be influenced by a high chloride mixture in the saliva or the intake of various foods and drinks with a low PH. Also, the characteristics of saliva change according to the patient's health and the time of day. We used mouthwashes in a static condition, but more metal release could occur in real life because of the fluidity of saliva in the mouth and also because oxide layers are removed by tooth brushing⁽²¹⁾.

CONCLUSION

1. Corrosion resistance of Dentos implants was higher compared to S.K surgical implants.
2. High amount of nickel ions were released in Listerine mouthwash as compared to Hexidine mouthwash
3. CHX-containing (HEXIDINE) mouthwash can be offered to patients who have orthodontic appliances rather than Na + alcohol-containing mouthwash (LISTERINE).

REFERENCES

1. Abraham ST, Paul MM. Micro implants for orthodontic anchorage: A review of complications and management. *J Dent Implant* 2013;3: 165-7.
2. Luzi C, Verna C, Melsen B 2007 A prospective clinical investigation of the failure rate of immediately loaded mini-implants used for orthodontic anchorage . *Progress in Orthodontics* 8 : 192 – 201.
3. Probst L , Lin W , Hutteman H 1992 Effect of fluoride prophylactic agents on titanium surfaces . *International Journal of Oral and Maxillofacial Implants* 7 : 390 – 394.
4. Costa M. Molecular mechanisms of nickel carcinogenesis. *Biol Chem* 2002;383:961-7.
5. Vreeburg KJ, de Groot K, von Blomberg M, Scheper RJ. Induction of immunological tolerance by oral administration of nickel and chromium. *J Dent Res* 1984;63:124-8.
6. Zhou D, Salnikow K, Costa M. Cap43, A novel gene specifically induced by Ni21 compounds. *Cancer Res* 1998;58:2182-9.
7. Schiff N, Grosogeat B, Lissac M, Dalard F. Influence of fluoridated mouthwashes on corrosion resistance of orthodontic wires. *Biomaterials* 2004;25:4535-42.
8. Shahla Momeni Danaei, Afsaneh Safavi, S. M. Mehdi Roeinpeikar, Morteza Oshagh, Shiva Iranpour, and Maryam Omidekhoda. Ion release from orthodontic brackets in 3 mouthwashes: An in-vitro study. *Am J Orthod Dentofacial Orthop* 2011;139:730-4.
9. Park HY, Shearer TR. In vitro release of nickel and chromium for simulated orthodontic appliances. *Am J Orthod* 1983;84:156-9.
10. Barrett RD, Bishara SE, Quinn JK. Biodegradation of orthodontic appliances. Part I. Biodegradation of nickel

and chromium in vitro . *Am J Orthod Dentofacial Orthop* 1993;103:8-14.

11. Hwang CJ, Shin JS, Cha JY. Metal release from simulated fixed orthodontic appliances. *Am J Orthod Dentofacial Orthop* 2001;120:383-91.
12. Schiff N, Boinet M, Morgon L, Lissac M, Dalard F, Grosogeat B. Galvanic corrosion between orthodontic wires and brackets in fluoride mouthwashes. *Eur J Orthod*. 2006;28:298–304.
13. Tuygun, Erdogan Ayse, Nalbantgil Didem, Ulkur Feyza, Sahin Fikrettin. 2014. Metal ion release from silver soldering and laser welding caused by different types of mouthwash. *The Angle orthodontist* doi:10.2319/050914-335.1.
14. Yokoyama K , Ichikawa T , Murakami H , Miyamoto Y , Asaoka K 2002 Fracture mechanisms of retrieved titanium screw thread in dental implant . *Biomaterials* 23 : 2459 – 2465.
15. Mockers O, Deroze D, Camps J. Cytotoxicity of orthodontic bands, brackets and archwires in vitro. *Dent Mater* 2002;18:311-7.
16. Tsaousidis G , Bauss O 2008 Influence of insertion site on the failure rates of orthodontic miniscrews. *Journal of Orofacial Orthopedics* 69 : 349 – 356.
17. Mouhyi J , Dohan Ehrenfest D M , Albrektsson T 2012 The peri-implantitis: implant surfaces, microstructure, and physicochemical aspects . *Clinical Implant Dentistry and Related Research* 14 : 170 – 183.
18. Schmalz G, Garhammer P. Biological interactions of dental cast alloys with oral tissues. *Dent Mater* 2002;18:396-406.
19. Rickles NH. Allergy in surface lesions of the oral mucosa. *Oral Surg Oral Med Oral Pathol* 1972;33:744-54.
20. Schriver WR, Shereff RH, Domnitz JM, Swintak EF, Civjan S. Allergic response to stainless steel wire. *Oral Surg Oral Med Oral Pathol* 1976;42:578-81.
21. Kerosuo H, Moe G, Kelven E. In vitro release of nickel and chromium from different types of simulated orthodontic appliances. *Angle Orthod* 1995;65:111-6.